



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

JRI-TMI USA, Inc.
11491 Sunset Hills Road, Suite 310
Reston, VA 20190

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 26 June 2028

Certificate Number: AC-2881



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

JRI-TMI USA, Inc.

11491 Sunset Hills Road, Suite 310
Reston, VA 20190

Guillaume Favre 703-668-0114

CALIBRATION

ISO/IEC 17025 Accreditation Granted: **24 June 2026**

Certificate Number: **AC-2881**

Certificate Expiry Date: **26 June 2028**


Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Absolute Pressure – Measuring Equipment	10 mbar to 5 bar 10 mbar to 15 bar 10 mbar to 30 bar	0.93 mbar 2.4 mbar 5.5 mbar	Comparison method with Mensor Pressure Controller CPC 6000/CPC 6050

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature– Measuring Equipment	(-80 to 140) °C	0.048 °C	Comparison method with Fluke Thermometer System 1502A-5628 and Fluke Baths 7321/7381
Temperature– Measuring Equipment	(30 to 350) °C	0.21 °C	Comparison method with Fluke Thermometer System 1502A-5628 and Fluke Metrology Well 9172
Relative Humidity– Measuring Equipment	(10 to 95) %RH	0.58 %RH	Comparison method with Thunder Scientific Humidity Generator 2900

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.



Jason Stine, Vice President

